

VERIFICATION OF CALIBRATION REPORT

JUL 28 2010

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130493</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>PATRICK SMITH</u>	ID# <u>6397</u> ✓ Date <u>28 JUL 10</u> ✓
A Agency <u>CLEAR SECURITY FORCES</u> Phone # <u>907-585-6293</u>	
Instrument Location <u>BLDG 37 CLEAR AFS, AK 99704</u>	
B Alco S/N <u>X124609</u> ✓ Target Value <u>-081</u> ✓ High Pressure <u>1125</u>	
Alco Test Values <u>-075</u> ✓ 1st Alco <u>-076</u> ✓ 2nd Alco	
Signature <u>[Signature]</u> CAB 7/29/10	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of AUGUST, 2010.

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JULY 22, 2010

OPERATOR'S NAME:

SMITH/PATRICK/J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 1234567

DEPT/AGENCY: CLR4

CASE/REPORT: VOC TEST

TEST TYPE: V

ALCO TARGET VALUE: .001

ALCO S/N: X124609

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.51 in

ALCO TARGET .079 08:44

BLANK TEST .000 08:45

INTERNAL STANDARD VERIFIED 08:45

ALCO TV 29.49 in .075 08:45

BLANK TEST .000 08:46

SUBJECT SAMPLE .000 08:47

BLANK TEST .000 08:47

ALCO TV 29.49 in .076 08:48

BLANK TEST .000 08:48

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JULY 22, 2010

TIME 08:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 29.49 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~